



ZIMMER BIOMET
Your progress. Our promise.®



Thoracolumbar Solutions

Polaris™ 5.5 Spinal System

Surgical Technique Guide




TABLE OF CONTENTS

Polaris 5.5 Spinal System Overview	4
Surgical Approach and Preparation	6
Pedicle Preparation	7
Screw Selection and Insertion	9
Screw Height Adjustment	11
Rod Application	11
Helical Flange® Plug Application	12
Final Locking	12
Surgical Options	
Reduction Options	14
Distraction and Compression	16
Cross Connector Application	17
Lateral Connectors	18
Closure	19
Implant Removal	19
Implants	20
Instrumentation	22
Recommended Ordering Guide	25
Kit Contents	26
Important Information on the Polaris 5.5 Spinal System	30

Zimmer Biomet Spine does not practice medicine. This technique was developed in conjunction with health care professionals. This document is intended for surgeons and is not intended for laypersons. Each surgeon should exercise his or her own independent judgment in the diagnosis and treatment of an individual patient, and this information does not purport to replace the comprehensive training surgeons have received. As with all surgical procedures, the technique used in each case will depend on the surgeon's medical judgment as the best treatment for each patient. Results will vary based on health, weight, activity and other variables. Not all patients are candidates for this product and/or procedure.

POLARIS 5.5 SPINAL SYSTEM OVERVIEW

We've Got the Lock on Pedicule Screw Technology


- Incorporates Helical Flange® Technology that minimizes seat splay and cross threading
- Load sharing, top loading, low profile system

The Polaris 5.5 Spinal System was created to offer a streamlined lumbar fixation system that uses a superior locking mechanism.

The System incorporates Helical Flange Technology that minimizes seat-splay and cross threading. The forces are concentrated inward, thus enabling the seat and plug to create a reliable mechanical lock.

The Polaris 5.5 Spinal System is a load-sharing, top-loading, low-profile system. The seat enables secure interface with the instruments for maximal manipulation agility. The design goals were to aid the surgeon with intraoperative efficiency and effectiveness while maintaining integrity and ease.

The Polaris 5.5 Spinal System is designed to address degenerative pathologies. The trays are configured to include multi-axial screws, extended screws, precut precontoured rods, Crossbar™ cross connectors*, lateral connectors and ergonomic instrumentation for maximum tactile feedback. The Polaris System continues to advance the spinal fixation needs of the aging population by providing fixation, variability and ease of use.




*The Crossbar™ cross connector was developed by SeaSpine, Inc.

System Design Features and Benefits



FEATURES	BENEFITS
Helical Flange Technology	Starts easily Minimizes cross threading and seat splay Forces are concentrated inward
5.5 mm rod system	Low profile Anatomic fit
Color-coded implants	Easy determination of screw sizes and instruments
Streamlined instruments	Logical and ergonomic
Screwdrivers	One piece, easy to use Provide for maximal visualization Control of screw trajectory
Minimum number of trays in operating room	Only two trays necessary per case
High on head instrument interface	Easy rod manipulation and excellent interface with instruments
Rod reducer	Very powerful and controlled User-friendly Locks onto seat for easy rod reduction
Multi-axial screws	60° of conical angulation for optimum versatility
Friction fit seat	Once the seat is in place, it remains in place for ease of rod introduction
Connection to Zimmer Biomet Spine posterior systems	Easy addition to the Altius M-INI system
Extended seat multi-axial screws	Allows for reducing a spondylolisthesis
Hydroxyapatite (HA) coated multi-axial screws	Dual lead screw thread for faster screw placement Balanced-start Tip geometry provides immediate engagement and minimizes toggle during insertion
Translation™ screw technology	3 mm of medial-lateral translation relative to the screw shaft Encourages optimal screw placement Less rod manipulation, easier rod introduction

SURGICAL APPROACH AND PREPARATION



STEP 1

- The patient is positioned prone as is customary for the surgeon, the spine is exposed subperiostally through a midline or paramedian incision, and a decompression is performed if indicated.

Note: Decortication must be performed meticulously.

Note: Graft can be placed or packed into the posterolateral gutters before or after the Polaris 5.5 Spinal System has been implanted.

PEDICLE PREPARATION



Open the entrance to the pedicle with the pedicle awl.



Prepare the pedicle hole with the reamer probe.



Prepare the pedicle hole with the chosen tap.

STEP 2

- After adequate exposure is achieved, the appropriate pedicle entry point is selected and the entrance to the pedicle is opened with an awl, burr or curette.
- The appropriate diameter reamer probe is used to prepare the pedicle using a slow circular motion, allowing the reamer probe to center itself along the longitudinal axis of the pedicle.
- Each reamer probe is marked with the major diameter of the screw with which it is to be used. The reamer probe is initially advanced to a depth of approximately 30 mm using the depth markings as a guide.

Note: Instead of a reamer probe, a pedicle probe can be used. The pedicle probe is used to create the pedicle hole by advancing the probe to a depth of approximately 30–40 mm using the depth markings as a guide.

STEP 3

- Although the screws are self-tapping, taps are available with the system and can be used to prepare the pedicle hole. Select the corresponding tap for the chosen screw diameter and advance the tap into the pedicle hole using the quick-connect handle.

PEDICLE PREPARATION (*Continued*)



Confirm containment of the pedicle with the pedicle sound.



Use the trial pins to ensure proper orientation and trajectory.

STEP 4

- The pedicle sound is then used to confirm bony containment of the pedicle hole by palpating all four walls and the bottom of the hole through the pedicle and into the vertebral body.
- The trial pins can be used to confirm proper orientation and trajectory.

SCREW SELECTION AND INSERTION



Select the appropriate screw size.

STEP 5

- Attach the multi-axial screwdriver to the quick connect handle by pulling back on the plunger at the base of the quick connect mechanism, inserting the shaft and releasing the plunger to lock the shaft in place.

Note: Self-tapping screws are available in several diameters and lengths.

Note: The appropriate screw length is determined using the depth markings on the pedicle probe or reamer probe.



Push the button located at the top of the knurled T, and load the screw onto multi-axial screwdriver.

STEP 6

- Hold the screw by the screw shaft and load the screw onto the tip of the multi-axial screwdriver.
- **Note:** The multi-axial screws may be loaded freehand or while seated within the surgical tray.
- Ensure that the male pentalobe at the distal tip of the multi-axial driver is fully seated within the female pentalobe located at the top of the screw shaft.
- Push the button located at the knurled T.
- Next, turn the knurled T in a clockwise direction to thread the outer shaft into the seat. Upon fully loading, the button on the T will release.
- Confirm that the screw is straight and secure in the driver.

SCREW SELECTION AND INSERTION (*Continued*)



Insert the screw into the pedicle.



Once inserted to desired depth, push the button in.



Turn the knurled T counterclockwise to release from the screw.

STEP 7

- The screw is advanced into the pedicle to the desired depth.
- During insertion, guide the driver by holding the blue sleeve on the shaft of the instrument.

STEP 8

- The driver is disengaged from the screw by pushing the button on the knurled T in and rotating the knurled T in a counterclockwise direction and then lifting the driver from the screw.

SCREW HEIGHT ADJUSTMENT



Use the dorsal height adjuster to adjust the screws.

STEP 9

- The dorsal height adjuster may be used to adjust the multi-axial screw height prior to rod placement.
- Seat the male pentalobe of the dorsal height adjuster into the female pentalobe located at the top of the screw shaft.
- Turn the adjuster for minor manipulation of the screw height.

ROD APPLICATION



Select appropriate length rod.



Set the dial on the rod bender to achieve the desired curvature.



Measure length of the rod using the rod template.



Insert rod using the rod holder.

STEP 10

- Once all screws have been inserted, the appropriate length rod should be chosen according to the construct.
- The rod template may be used to aid in rod selection. The rod should project at least 2.0 mm beyond the screw seats at the end of the construct.
- Be sure to account for large curves and distractions when choosing rod length. If necessary, the selected rod may be contoured with the rod bender.

HELICAL FLANGE PLUG APPLICATION



Load plug onto the double end plug starter.



Insert plug.

STEP 11

- When all screws have been inserted and the rods have been placed in the screw seats, the construct is then secured using Helical Flange plugs.
- One plug is firmly pressed onto each end of the double end plug starter. All plugs should be placed and then provisionally tightened.
- If necessary, the plug starter may be used in combination with the rod persuader, reduction fork or rod pusher.

Note: Reference pages 14 and 15 for reduction options.

FINAL LOCKING

STEP 12

After provisional tightening, proper implant placement should be confirmed with radiographs. The plugs are then tightened with either the torque-indicating wrench or the torque-limiting wrench in combination with the torque stabilizer.

- Insert the chosen torquing device through the center of the torque stabilizer. Position the tip of the torque wrench into the plug. Seat the distal end of the torque stabilizer over the screw seat and confirm that the stabilizer fits firmly on the rod. The rod will be positioned within the slots of the stabilizer.



Arrows of the torque-indicating wrench align at 0, signifying the start position. When the torque level is achieved, the arrow will align at 110 in-lbs. **THERE IS NO AUDIBLE CLICK.**



Turn the torque-limiting wrench clockwise until an audible click is heard at 110 in-lbs of torque.

Final Tightening — Torque-indicating Handle

- The torque-indicating wrench is turned in a clockwise direction while the torque stabilizer is held with resistive force in a counterclockwise direction. Two etched arrows indicate when the appropriate torque is obtained. The first set of arrows lines up showing the start position at zero. Upon reaching the intended final torque, two arrows will line up at 110 in-lbs.

Note: *THERE IS NO AUDIBLE CLICK with the torque-indicating wrench. Over-torquing with the torque-indicating wrench (turning beyond the point where the arrows line up) may damage the wrench. Always ensure the wrench indicates 0 in-lbs. of torque prior to use.*

Final Tightening — Torque-limiting Handle

- The torque-limiting handle attaches to the plug driver. The torque-limiting wrench is turned in a clockwise direction while the torque stabilizer is held with resistive force in a counterclockwise direction. The torque-limiting wrench should be turned until an audible click is heard, applying 110 in-lbs of torque.

Note: *Use the chosen torque instrument in combination with the torque stabilizer.*

SURGICAL OPTIONS

Reduction Options



The persuader may be used to fully seat the rod in the screw seat.




Reduction fork

Position the reduction fork under the screw seat and tilt the instrument to persuade the rod into the screw seat.

STEP 12 (Continued)


- When using the rod persuader, place the persuader over the top of the screw seat. The internal stop of the persuader will ensure that the instrument is in the correct position on the seat to facilitate manipulation.
- Squeeze the handle of the rod persuader to fully seat the rod in the screw seat. The plug starter will fit through the cannulated portion of the persuader, allowing for plug application with the rod persuader in place.
- To release the persuader, press the trigger located underneath the handle. Once released, the persuader may then be removed from the screw seat.
- When using the reduction fork, position the fork section underneath screw seat. Tilt the reduction fork to persuade the rod into the screw seat.



Push the rod down to persuade the rod into the seat and insert the plug.



Torque stabilizer may be used to guide the plug starter.



The soft tissue retractor aids retraction of the soft tissue away from the screw seat.

- When using the rod pusher, place the distal tip onto the rod and push the rod down to persuade the rod into the screw seat.
- The torque stabilizer may be used to reposition the axis of the screw seat while simultaneously acting as a guide for the plug starter.

Note: *If soft tissue is interfering with proper plug placement, the soft tissue retractor may be used to retract the soft tissue away from the screw by placing the bifid tip of the retractor under the screw seat.*

SURGICAL OPTIONS (Continued)

Distraction and Compression



Provisionally tighten the plug while the compressor or distractor is in place.

- Distraction and compression can be achieved by using either the distractor or compressor. Both instruments permit intraoperative application of linear distraction or compression at any level.
- The distal tips of the distractor or compressor are placed on the rod, and the desired degree of distraction or compression is applied.
- The distraction or compression device will maintain the position of the vertebra until the plug is provisionally tightened with the plug starter connected to the chosen quick connect handle.

Cross Connector Application



Select the appropriate sized cross connector.



Torque the set screws on the cross connector. Torque until an audible click is heard to apply 40 in-lbs.

-
- In the event that additional torsional stability is required, a cross connector can be used. The cross connector should be applied after the construct has been assembled and the plugs have been tightened.
 - Apply the cross connector to the rods and tighten the screws with the cross connector torque wrench until an audible click is heard, applying 40 in-lbs of torque to the set screws (tighten the outer set screws, then the central set screw).

SURGICAL OPTIONS (Continued)

Lateral Connectors



Lateral connector — 25 mm (length is measured from center of the seat to the end of the rod).



The lateral connector is applied to the lateral screw at L5.



The lateral connectors are secured with the same Helical Flange plugs as the screws.

- Lateral connectors can be used if screw placement requires a severe bend in the rod. The lateral connectors allow for an offset, thus minimizing rod bending. The lateral connectors are secured with the same Helical Flange plug as the pedicle screws.
- Place the arm of the lateral connector in the pedicle screw seat and secure the lateral connector in place by provisionally tightening the plug.
- Place the longitudinal rod into the seat of the lateral connector. Once the rod has been placed, insert the Helical Flange plug into the seat of the lateral connector (refer to “Helical Flange Plug Application” and “Final Locking”).

CLOSURE

IMPLANT REMOVAL



Use the multi-axial screwdriver to remove the screw.

-
- After implantation of the Polaris 5.5 Spinal System is complete, wound closure is performed according to the standard protocol for the surgeon.

-
- The multi-axial screwdriver is used to remove the multi-axial screws by seating the male pentalobe end into the female pentalobe at the top of the screw shaft.
 - Slide the outer sleeve down and turn the large knurled T clockwise to lock into the screw seat.
 - Once the driver is tightened, the screw may be backed out of the pedicle.
 - Removal of the Polaris 5.5 Spinal System is performed by reversing the order of the implant procedure. The quick connect fixed T-handle attached to the plug driver in combination with the torque stabilizer must be used first to remove the plugs.

Note: When removing previously torqued plugs, turn the fixed T-handle in a slight clockwise direction before turning counterclockwise. Continue with this back-and-forth motion until the plug loosens.


IMPLANTS



Helical Flange Plug

PART NUMBER

2000-1005



Lateral Connector, 25 mm

PART NUMBER

2000-1020



Multi-axial Screws

Multi-axial Screws are available in 4.75, 5.5, 6.5, 7.5 and 8.5 mm diameters in 30–55 mm lengths.



Multi-axial Reduction Screws

Multi-axial Reduction Screws are available in 5.5, 6.5 and 7.5 mm diameter in 30–55 mm lengths.



Translation Screws

Translation Screws are available in 4.75, 5.5, 6.5 and 7.5 mm diameters in 30–55 mm lengths.



Translation Iliac Screws

Translation Iliac Screws are available in 6.5, 7.5 and 8.5 mm diameters in 35–100 mm lengths.



HA Coated Screws

Hydroxyapatite (HA) Coated Screws are available in 5.5, 6.5, 7.5 and 8.5 mm diameters in a variety of lengths.

**Pre-cut, Contoured Rods**

PART NUMBER

Ti Alloy 30–100 mm, increments of 5 mm

See pages 27, 28

**Telescoping and Angulating
Crossbar Cross Connectors**

PART NUMBER

Ranging in length from 16–75 mm

See pages 27, 28

INSTRUMENTATION



Fixed T-handle PART NUMBER
95697



Fixed Handle–Straight PART NUMBER
94699



Ratchet Handle–Straight PART NUMBER
124799



Ratchet T-handle PART NUMBER
124797



Fixed Teardrop Handle PART NUMBER
2000-9006



Ratcheting Teardrop Handle PART NUMBER
2000-6481



Awl Shaft PART NUMBER
94505



Thoracic Pedicle Probe PART NUMBER
14-500100



Straight Pedicle Probe PART NUMBER
14-500101



Curved Pedicle Probe PART NUMBER
14-500102



Reamer Probe PART NUMBER

4.75 mm	2000-9091
5.5 mm	2000-9092
6.5 mm	2000-9093
7.5 mm	2000-9094



Tap PART NUMBER

4.75 mm	2000-9023
5.5 mm	2000-9024
6.5 mm	2000-9025
7.5 mm	2000-9026
8.6 mm	2000-9027



Stiff Pedicle Sound PART NUMBER
4010



Flexible Pedicle Sound PART NUMBER
2000-9015



Trial Pins PART NUMBER
9 cm 4077
11 cm 4072



Screwdriver PART NUMBER
Multi-axial Screw Inserter 14-500185
Translation Screwdriver 14-500070



Dorsal Height Adjuster PART NUMBER
2000-9072



Rod Template PART NUMBER
94612



Rod Bender PART NUMBER
2000-9044



Rod Holder PART NUMBER
94613



Plug Driver PART NUMBER
2000-9061



Double End Plug Starter PART NUMBER
2000-9060



Straight Rod Pusher PART NUMBER
2000-9059



Reduction Fork PART NUMBER
2000-9054



Rod Persuader PART NUMBER
Parallel 2000-9055
Perpendicular 14-500198



Soft Tissue Retractor PART NUMBER
94614

INSTRUMENTATION (Continued)

**Compressor**

PART NUMBER

94686

**Distractor**

PART NUMBER

94687

**Cross Connector Torque Wrench**

PART NUMBER

94624

**Torque-limiting Wrench**

PART NUMBER

94522

**Torque-indicating Wrench**

PART NUMBER

2000-9082

**Torque Stabilizer**

PART NUMBER

2000-9075

**Reduction Screw Break-off Stabilizer**

PART NUMBER

2000-9019

**Reduction Screw Break-off Plier**

PART NUMBER

2000-9074

**Short Rocket Threaded Persuader**

PART NUMBER

14-500200

**Offset T-handle**

PART NUMBER

14-500201

**T-handle**

PART NUMBER

14-500203

**Quick-connect Adaptor**

PART NUMBER

14-500202

Recommended Ordering Guide: Polaris 5.5 Degenerative

REQUIRED KITS

Instrument Kits

DESCRIPTION	KIT NUMBER
Standard Instrument Kit	55500146
OR	
Standard Instrument Kit A	14-509680
Standard Instrument Kit B	14-509681

Implant Kits

DESCRIPTION	KIT NUMBER
Standard Implant Kit	55500147


OPTIONAL AUXILIARY KITS

Instrument Kits

DESCRIPTION	KIT NUMBER
Short Rocket Threaded Persuader Kit	14-509639
Perpendicular Persuader Kit	14-509637

Implant Kits

DESCRIPTION	KIT NUMBER
Multi-axial Reduction Screw Implant Kit	14-509605
4.75 Multi-axial Screw Implant Kit	14-509606
8.5 Multi-axial Screw Implant Kit	14-509607
HA Coated Multi-axial Screw Implant Kit	14-509700
Iliac Screw Implant Kit	14-509635
Translation Screw Standard Implant Kit	14-509669
Translation 4.75 Screw Implant Kit	14-509682
Translation Iliac Screw Implant Kit	14-509668



KIT CONTENTS

Standard Instrument Kit Kit Number: 55500146

DESCRIPTION	QTY	PART NUMBER
Plug Driver	2	2000-9061
Multi-axial Screw Insertter	2	14-500185
Torque Stabilizer	1	2000-9075
Torque-limiting Wrench, 110 in-lbs	1	94522
Awl Shaft	1	94505
Thoracic Pedicle Probe	1	14-500100
Straight Pedicle Probe	1	14-500101
Curved Pedicle Probe	1	14-500102
Flexible Pedicle Sound	1	2000-9015
Stiff Pedicle Sound	1	4010
9.0 cm Trial Pin	4	4077
11 cm Trial Pin	4	4072
4.75 mm Tap	1	2000-9023
5.5 mm Tap	1	2000-9024
6.5 mm Tap	1	2000-9025
7.5 mm Tap	1	2000-9026
8.5 mm Tap	1	2000-9027
Reduction Fork	1	2000-9054
Rod Persuader	1	2000-9055
Straight Rod Pusher	1	2000-9059
Double End Plug Starter	2	2000-9060
Dorsal Height Adjuster	1	2000-9072
Reduction Screw Break-off Plier	1	2000-9074
4.75 mm Reamer Probe	1	2000-9091
5.5 mm Reamer Probe	1	2000-9092
6.5 mm Reamer Probe	1	2000-9093
7.5 mm Reamer Probe	1	2000-9094
Ratchet T-handle	1	124797
Ratchet Handle–Straight	1	124799
Rod template	1	94612
Rod Holder	1	94613
Soft Tissue Retractor	1	94614
Cross Connector Torque Wrench	1	94624
Parallel Compressor	1	94686
Parallel Distractor	1	94687
Rod Bender	1	2000-9044
Fixed T-handle	1	94697
Fixed Handle–Straight	1	94699
Fixed Teardrop Handle	1	2000-9006
Reduction Screw Break-off stabilizer	1	2000-9019

Standard Instrument Kit A Kit Number: 14-509680

DESCRIPTION	QTY	PART NUMBER
Plug Driver	2	2000-9061
Multi-axial Screw Insertter	2	14-500185
Rod Rocker, Extended Throw	1	14-500197
Torque Stabilizer	1	14-500018
Torque-limiting Wrench, 110 in-lbs	2	94522
Awl Shaft	1	94505
Thoracic Pedicle Probe	1	14-500001
Straight Pedicle Probe	1	14-500002
Curved Pedicle Probe	1	14-500003
Firm Pedicle Sound	1	14-500007
Flexible Pedicle Sound	1	2000-9015
4.75 mm Tap	1	2000-9023
5.5 mm Tap	1	2000-9024
6.5 mm Tap	1	2000-9025
7.5 mm Tap	1	2000-9026
8.5 mm Tap	1	2000-9027
Screw Head Positioner	1	14-500072
Plug Starter	2	14-500170
Dorsal Height Adjuster	1	14-501680
Ratchet T-handle	1	124797
Ratchet Handle–Straight	1	124799
Rod Holder	1	94613
Cross Connector Torque Wrench	1	94624
Ratcheting Teardrop Handle	1	2000-6481

Standard Instrument Kit B Kit Number: 14-509681

DESCRIPTION	QTY	PART NUMBER
Reduction Screw Break-off Plier	1	14-500009
Perpendicular Rod Persuader	1	14-500198
Rod Bender	1	2000-9044
Parallel Compressor	1	94686
Parallel Distractor	1	94687

Short Rocket Threaded Persuader Kit Kit Number: 14-509639

DESCRIPTION	QTY	PART NUMBER
Short Threaded Rod Persuader	6	14-500200
T-handle Persuader	2	14-500201
Quick-connect Adaptor	1	14-500202

Perpendicular Rod Persuader Kit Kit Number: 14-509637

DESCRIPTION	QTY	PART NUMBER
Perpendicular Rod Persuader	2	14-500198

Hydroxyapatite (HA) Coated Multi-axial Screw Kit Kit Number: 14-509700*

DESCRIPTION	QTY	PART NUMBER
ø5.5 mm × 30 mm HA Multi-axial Screw	4	14-592330
ø5.5 mm × 35 mm HA Multi-axial Screw	6	14-592335
ø5.5 mm × 40 mm HA Multi-axial Screw	8	14-592340
ø5.5 mm × 45 mm HA Multi-axial Screw	8	14-592345
ø5.5 mm × 50 mm HA Multi-axial Screw	4	14-592350
ø6.5 mm × 35 mm HA Multi-axial Screw	6	14-592435
ø6.5 mm × 40 mm HA Multi-axial Screw	8	14-592440
ø6.5 mm × 45 mm HA Multi-axial Screw	8	14-592445
ø6.5 mm × 50 mm HA Multi-axial Screw	8	14-592450
ø6.5 mm × 55 mm HA Multi-axial Screw	6	14-592455
ø7.5 mm × 40 mm HA Multi-axial Screw	8	14-592540
ø7.5 mm × 45 mm HA Multi-axial Screw	8	14-592545
ø7.5 mm × 50 mm HA Multi-axial Screw	6	14-592550
ø7.5 mm × 55 mm HA Multi-axial Screw	6	14-592555
ø8.5 mm × 40 mm HA Multi-axial Screw	8	14-592640
ø8.5 mm × 45 mm HA Multi-axial Screw	8	14-592645
ø8.5 mm × 50 mm HA Multi-axial Screw	8	14-592650
ø8.5 mm × 55 mm HA Multi-axial Screw	6	14-592655

*HA screws have a dual lead thread. Double-lead taps from the Translation Implant Kit (14-509669) must be ordered separately.

Standard Implant Kit Kit Number: 55500147

DESCRIPTION	QTY	PART NUMBER
Plug	30	2000-1005
Lateral Connector, 25 mm	2	2000-1020
ø5.5 mm × 30 mm Multi-axial Screw	6	2000-2330
ø5.5 mm × 35 mm Multi-axial Screw	6	2000-2335
ø5.5 mm × 40 mm Multi-axial Screw	6	2000-2340
ø5.5 mm × 45 mm Multi-axial Screw	6	2000-2345
ø5.5 mm × 50 mm Multi-axial Screw	6	2000-2350
ø5.5 mm × 55 mm Multi-axial Screw	4	2000-2355
ø6.5 mm × 30 mm Multi-axial Screw	4	2000-2430
ø6.5 mm × 35 mm Multi-axial Screw	6	2000-2435
ø6.5 mm × 40 mm Multi-axial Screw	8	2000-2440
ø6.5 mm × 45 mm Multi-axial Screw	8	2000-2445
ø6.5 mm × 50 mm Multi-axial Screw	8	2000-2450
ø6.5 mm × 55 mm Multi-axial Screw	4	2000-2455
ø7.5 mm × 30 mm Multi-axial Screw	4	2000-2530
ø7.5 mm × 35 mm Multi-axial Screw	4	2000-2535
ø7.5 mm × 40 mm Multi-axial Screw	6	2000-2540
ø7.5 mm × 45 mm Multi-axial Screw	6	2000-2545
ø7.5 mm × 50 mm Multi-axial Screw	6	2000-2550
ø7.5 mm × 55 mm Multi-axial Screw	4	2000-2555
30 mm Ti Alloy Curved Rod	4	2000-5130
35 mm Ti Alloy Curved Rod	4	2000-5135
40 mm Ti Alloy Curved Rod	4	2000-5140
45 mm Ti Alloy Curved Rod	4	2000-5145
50 mm Ti Alloy Curved Rod	4	2000-5150
55 mm Ti Alloy Curved Rod	4	2000-5155
60 mm Ti Alloy Curved Rod	4	2000-5160
65 mm Ti Alloy Curved Rod	4	2000-5165
70 mm Ti Alloy Curved Rod	4	2000-5170
75 mm Ti Alloy Curved Rod	4	2000-5175
80 mm Ti Alloy Curved Rod	4	2000-5180
90 mm Ti Alloy Curved Rod	4	2000-5190
100 mm Ti Alloy Curved Rod	4	2000-5199
510 mm Ti Alloy Rod (with Hex)	2	2000-5405
XXSmall Cross Connector	2	94669
XSmall Cross Connector	2	94670
Small Cross Connector	2	94671
Medium Cross Connector	2	94672
Large Cross Connector	2	94673

KIT CONTENTS (Continued)

4.75 mm Multi-axial Screw Implant Kit Kit Number: 14-509606

DESCRIPTION	QTY	PART NUMBER
ø4.75 mm × 20 mm Multi-axial Screw	12	2000-2220
ø4.75 mm × 25 mm Multi-axial Screw	12	2000-2225
ø4.75 mm × 30 mm Multi-axial Screw	12	2000-2230
ø4.75 mm × 35 mm Multi-axial Screw	12	2000-2235
ø4.75 mm × 40 mm Multi-axial Screw	12	2000-2240
ø4.75 mm × 45 mm Multi-axial Screw	6	2000-2245
ø4.75 mm × 50 mm Multi-axial Screw	6	2000-2250

4.75 mm Translation Screw Implant Kit Kit Number: 14-509682

DESCRIPTION	QTY	PART NUMBER
ø4.75 mm × 25 mm Translation Screw	4	14-578225
ø4.75 mm × 30 mm Translation Screw	6	14-578230
ø4.75 mm × 35 mm Translation Screw	6	14-578235
ø4.75 mm × 40 mm Translation Screw	6	14-578240
ø4.75 mm × 45 mm Translation Screw	2	14-578245

8.5 mm Multi-axial Screw Implant Kit Kit Number: 14-509607

DESCRIPTION	QTY	PART NUMBER
ø8.5 mm × 30 mm Multi-axial Screw	4	2000-2630
ø8.5 mm × 35 mm Multi-axial Screw	4	2000-2635
ø8.5 mm × 40 mm Multi-axial Screw	4	2000-2640
ø8.5 mm × 45 mm Multi-axial Screw	4	2000-2645
ø8.5 mm × 50 mm Multi-axial Screw	4	2000-2650
ø8.5 mm × 55 mm Multi-axial Screw	4	2000-2655

Translation Screw Implant Kit Kit Number: 14-509669

DESCRIPTION	QTY	PART NUMBER
Button Lock Screw Inserter	2	14-500070
Translating Screw Bone Planer	1	14-500071
Double-lead Tap, 4 mm	1	14-500085
Double-lead Tap, 4.75 mm	1	14-500086
Double-lead Tap, 5.5 mm	1	14-500088
Double-lead Tap, 6.5 mm	1	14-500089
Double-lead Tap, 7.5 mm	1	14-500090

Translation Screw Implant Kit Kit Number: 14-509669 (Continued)

DESCRIPTION	QTY	PART NUMBER
ø5.5 mm × 25 mm Translation Screw	2	14-578325
ø5.5 mm × 30 mm Translation Screw	6	14-578330
ø5.5 mm × 35 mm Translation Screw	8	14-578335
ø5.5 mm × 40 mm Translation Screw	8	14-578340
ø5.5 mm × 45 mm Translation Screw	8	14-578345
ø5.5 mm × 50 mm Translation Screw	4	14-578350
ø5.5 mm × 55 mm Translation Screw	2	14-578355
ø6.5 mm × 30 mm Translation Screw	4	14-578430
ø6.5 mm × 35 mm Translation Screw	6	14-578435
ø6.5 mm × 40 mm Translation Screw	8	14-578440
ø6.5 mm × 45 mm Translation Screw	8	14-578445
ø6.5 mm × 50 mm Translation Screw	8	14-578450
ø6.5 mm × 55 mm Translation Screw	6	14-578455
ø7.5 mm × 30 mm Translation Screw	4	14-578530
ø7.5 mm × 35 mm Translation Screw	6	14-578535
ø7.5 mm × 40 mm Translation Screw	8	14-578540
ø7.5 mm × 45 mm Translation Screw	8	14-578545
ø7.5 mm × 50 mm Translation Screw	8	14-578550
ø7.5 mm × 55 mm Translation Screw	6	14-578555
Plug	30	2000-1005
Lateral Connector, Open, 25 mm	2	2000-1020
30 mm Ti Alloy Curved Rod	4	2000-5130
35 mm Ti Alloy Curved Rod	4	2000-5135
40 mm Ti Alloy Curved Rod	4	2000-5140
45 mm Ti Alloy Curved Rod	4	2000-5145
50 mm Ti Alloy Curved Rod	4	2000-5150
55 mm Ti Alloy Curved Rod	4	2000-5155
60 mm Ti Alloy Curved Rod	4	2000-5160
65 mm Ti Alloy Curved Rod	4	2000-5165
70 mm Ti Alloy Curved Rod	4	2000-5170
75 mm Ti Alloy Curved Rod	4	2000-5175
80 mm Ti Alloy Curved Rod	4	2000-5180
90 mm Ti Alloy Curved Rod	4	2000-5190
100 mm Ti Alloy Curved Rod	4	2000-5199
510 mm Ti Alloy Rod with Hex	2	2000-5405
XXSmall Cross Connector	2	94669
XSmall Cross Connector	2	94670
Small Cross Connector	2	94671
Medium Cross Connector	2	94672
Large Cross Connector	2	94673

Multi-axial Reduction Screw Implant Kit Kit Number: 14-509605

DESCRIPTION	QTY	PART NUMBER
∅5.5 mm × 30 mm Multi-axial Reduction Screw	4	2000-7330
∅5.5 mm × 35 mm Multi-axial Reduction Screw	4	2000-7335
∅5.5 mm × 40 mm Multi-axial Reduction Screw	4	2000-7340
∅5.5 mm × 45 mm Multi-axial Reduction Screw	4	2000-7345
∅5.5 mm × 50 mm Multi-axial Reduction Screw	4	2000-7350
∅5.5 mm × 55 mm Multi-axial Reduction Screw	2	2000-7355
∅6.5 mm × 30 mm Multi-axial Reduction Screw	4	2000-7430
∅6.5 mm × 35 mm Multi-axial Reduction Screw	6	2000-7435
∅6.5 mm × 40 mm Multi-axial Reduction Screw	8	2000-7440
∅6.5 mm × 45 mm Multi-axial Reduction Screw	8	2000-7445
∅6.5 mm × 50 mm Multi-axial Reduction Screw	6	2000-7450
∅6.5 mm × 55 mm Multi-axial Reduction Screw	4	2000-7455
∅7.5 mm × 30 mm Multi-axial Reduction Screw	2	2000-7530
∅7.5 mm × 35 mm Multi-axial Reduction Screw	6	2000-7535
∅7.5 mm × 40 mm Multi-axial Reduction Screw	6	2000-7540
∅7.5 mm × 45 mm Multi-axial Reduction Screw	6	2000-7545
∅7.5 mm × 50 mm Multi-axial Reduction Screw	4	2000-7550
∅7.5 mm × 55 mm Multi-axial Reduction Screw	2	2000-7555

Multi-axial Iliac Screw Implant Kit Kit Number: 14-509635

DESCRIPTION	QTY	PART NUMBER
∅6.5 mm × 60 mm Multi-axial Iliac Screw	4	14-500290
∅6.5 mm × 70 mm Multi-axial Iliac Screw	4	14-500292
∅6.5 mm × 80 mm Multi-axial Iliac Screw	4	14-500294
∅6.5 mm × 90 mm Multi-axial Iliac Screw	4	14-500296
∅7.5 mm × 60 mm Multi-axial Iliac Screw	4	14-500310
∅7.5 mm × 70 mm Multi-axial Iliac Screw	4	14-500312
∅7.5 mm × 80 mm Multi-axial Iliac Screw	4	14-500314
∅7.5 mm × 90 mm Multi-axial Iliac Screw	4	14-500316
∅8.5 mm × 60 mm Multi-axial Iliac Screw	4	14-500330
∅8.5 mm × 70 mm Multi-axial Iliac Screw	4	14-500332
∅8.5 mm × 80 mm Multi-axial Iliac Screw	4	14-500334
∅8.5 mm × 90 mm Multi-axial Iliac Screw	4	14-500336

Translation Screw Iliac Implant Kit Kit Number: 14-509668

DESCRIPTION	QTY	PART NUMBER
Screw Shaft Remover	1	14-500073
Double-lead Iliac Tap, 5.5 mm	1	14-500191
Double-lead Iliac Tap, 6.5 mm	1	14-500192
Double-lead Iliac Tap, 7.5 mm	1	14-500193
Double-lead Iliac Tap, 8.5 mm	1	14-500194
Double-lead Iliac Tap, 9.5 mm	1	14-500097
Double-lead Iliac Tap, 10.5 mm	1	14-500098
∅6.5 mm × 60 mm Iliac Screw	4	14-578460
∅6.5 mm × 70 mm Iliac Screw	4	14-578470
∅6.5 mm × 80 mm Iliac Screw	2	14-578480
∅6.5 mm × 90 mm Iliac Screw	2	14-578490
∅7.5 mm × 60 mm Iliac Screw	4	14-578560
∅7.5 mm × 70 mm Iliac Screw	4	14-578570
∅7.5 mm × 80 mm Iliac Screw	4	14-578580
∅7.5 mm × 90 mm Iliac Screw	4	14-578590
∅7.5 mm × 100 mm Iliac Screw	2	14-578599
∅6.5 mm × 90 mm Iliac Screw	2	14-578490
∅7.5 mm × 60 mm Iliac Screw	4	14-578560
∅7.5 mm × 70 mm Iliac Screw	4	14-578570
∅7.5 mm × 80 mm Iliac Screw	4	14-578580
∅7.5 mm × 90 mm Iliac Screw	4	14-578590
∅7.5 mm × 100 mm Iliac Screw	2	14-578599
∅8.5 mm × 35 mm Iliac Screw	4	14-578635
∅8.5 mm × 40 mm Iliac Screw	4	14-578640
∅8.5 mm × 45 mm Iliac Screw	4	14-578645
∅8.5 mm × 50 mm Iliac Screw	4	14-578650
∅8.5 mm × 55 mm Iliac Screw	4	14-578655
∅8.5 mm × 60 mm Iliac Screw	4	14-578660
∅8.5 mm × 70 mm Iliac Screw	4	14-578670
∅8.5 mm × 80 mm Iliac Screw	4	14-578680
∅8.5 mm × 90 mm Iliac Screw	4	14-578690
∅8.5 mm × 100 mm Iliac Screw	2	14-578699

IMPORTANT INFORMATION ON THE POLARIS 5.5 SPINAL SYSTEM

Device Description

The Polaris Spinal System is a non-cervical spinal fixation system. The system includes screws, HA coated screws, various types and sizes of rods, locking nuts, hooks, lateral connectors, plugs, washers, staples, rod connectors/ dominos and various cross connectors. Various instruments are also available for use by the surgeon to facilitate implantation of the device.

Indications for Use

The Polaris Spinal System is a non-cervical spinal fixation device intended for immobilization and stabilization as an adjunct to fusion as a pedicle screw fixation system, a posterior hook and sacral/iliac screw fixation system, or as an anterior or anterolateral fixation system for use with autograft and/or allograft. The Polaris Spinal System is indicated for all the following conditions: degenerative disc disease (defined as discogenic back pain with degeneration of the disc confirmed by history and radiographic studies), spondylolisthesis, trauma, (i.e., fracture or dislocation), deformity or curvature (i.e., scoliosis, kyphosis, Scheuermann's disease, and/or lordosis), tumor, stenosis, pseudoarthrosis or failed previous fusion.

The Ballista® Instruments* are intended to be used with the Ballista/Polaris 5.5 implants. Cannulated screws and percutaneous rods may be used with the Ballista instruments to provide the surgeon with a percutaneous approach for posterior spinal surgery for the above indications.

For pediatric patients, the Polaris Spinal System may be used for posterior, non-cervical pedicle screw fixation as an adjunct to fusion to treat adolescent idiopathic scoliosis and is also indicated for treatment of the following conditions: spondylolisthesis/spondylolysis and fractures caused by tumor and/or trauma. Pedicle screw fixation is limited to a posterior approach.

The Polaris Spinal System may be used with the instruments in AccuVision® Minimally Invasive Spinal Exposure System* to provide the surgeon with a minimally invasive approach for posterior spinal surgery.

The dominos in the Polaris Spinal System can be used to connect the Polaris Spinal System to the Altius™ Spinal System*, the Lineum® OCT Spine System*, the Array® Spinal System*, the Biomet® Omega21™ Spinal System* or the Synergy™ Spinal System* to achieve additional levels of fixation.

Please refer to the individual system's Package Insert for a list of the indications for use for each system.

Contraindications

- Spinal infection
- Morbid obesity
- A patient who in the surgeon's opinion is not psychosocially, mentally or physically able to fully comply with the post-operative treatment regime (e.g., mental illness, alcoholism or drug abuse.)
- Pregnancy
- Metal sensitivity/foreign body sensitivity
- Patients with inadequate tissue coverage over the operative site
- Open wounds local to the operative area
- Direct current stimulation, when used in a construct that includes Translation Screws

Warnings

The safety and effectiveness of pedicle screw spinal systems have been established only for spinal conditions with significant mechanical instability or deformity requiring fusion with instrumentation. These conditions are significant mechanical instability or deformity of the thoracic, lumbar, and sacral spine secondary to severe spondylolisthesis (grades 3 and 4) of the L5–S1 vertebrae, degenerative spondylolisthesis with objective evidence of neurologic impairment, fracture, dislocation, scoliosis, kyphosis, spinal tumor and failed previous fusion (pseudarthrosis). The safety and effectiveness of these devices for any other conditions are unknown.

Potential risks identified with the use of this device, which may require additional surgery, include device component fracture, loss of fixation, nonunion, fracture of the vertebra, neurological injury and vascular or visceral injury.

Implant Strength and Loading. The Polaris Spinal System is intended to assist healing and is not intended to replace normal bony structures. Loads produced by weight bearing and activity levels will dictate the longevity of the implant. These devices are not designed to withstand the unsupported stress of full weight bearing or load bearing, and cannot withstand activity levels and/or loads equal to those placed on normal healthy bone. If healing is delayed or does not occur, the implant could eventually break because of metal fatigue

* These optional components are not approved in all regions.

Therefore, it is important that immobilization of the operative site be maintained until firm bony union (confirmed by clinical and radiographic examination) is established.

The surgeon must be thoroughly knowledgeable of the medical, surgical, mechanical and metallurgical aspects of the Polaris Spinal System. Postoperative care is extremely important.

The patient should be warned that noncompliance with postoperative instructions could lead to breakage of the implant and/or possible migration requiring revision surgery to remove the implant.

Selection of Implants. Selection of the proper size, shape and design of the implant increases the potential for success. While proper selection can help to minimize risks, the size and shape of human bones present limitations on the size and strength of implants.

Metabolic bone disease such as severe osteoporosis may adversely affect adequate fixation of the implants because of the poor quality of the bone

The surgeon must ensure that all necessary implants and instruments are on hand before surgery. They must be handled and stored carefully and protected from damage, including from corrosive environments. They should be carefully unpacked and inspected for damage before use. All nonsterile components and instruments must be cleaned and sterilized before use. Zimmer Biomet Spine implants should not be used with implants or instruments from another manufacturer for reasons of metallurgy, mechanics and design.

Corrosion. Contact of dissimilar metals accelerates the corrosion process, which could increase the possibility of fatigue fracture of the implants. Therefore, only use like or compatible metals for implants that are in contact with each other. Never use stainless steel and titanium implant components in the same construct. Cobalt Chrome Alloy rods should not be used with Stainless Steel Components. Cobalt Chrome Alloy rods are to be used ONLY with titanium implant components in the same construct.

The Polaris Spinal System has not been evaluated for safety and compatibility in the magnetic resonance environment and has not been tested for heating or migration in the MR environment.

The safety and effectiveness of this device has not been established for use as part of a growing rod construct. This device is intended to be used only when definitive fusion is being performed at all instrumented levels.

Please refer to the Package Insert and/or surgical technique for the proper use of these types of devices.

Precautions

Do not reuse implants/devices. While an implant/device may appear undamaged, previous stress may have created imperfections that would reduce the service life of the implant/device. Do not treat patients with implants/devices that have been even momentarily placed in or used on a different patient.

Handling of Implants. If contouring of the rod is required, avoid sharp bends and reverse bends. Avoid notching or scratching of the device, which could increase internal stresses and lead to early breakage.

Implant Removal After Healing. After healing is complete, the implant is intended to be removed since it is no longer necessary. Implants that are not removed may result in complications such as implant loosening, fracture, corrosion, migration, pain or stress shielding of bone, particularly in young, active patients. Implant removal should be followed by adequate postoperative management.

Adequate Patient Instructions. A patient must be instructed on the limitations of the metallic implant, and should be cautioned regarding physical activity and weight bearing or load bearing before complete healing.

Surgical Techniques. The implantation of pedicle screw spinal systems should be performed only by experienced spinal surgeons with specific training in the use of pedicle screw spinal systems because this is a technically demanding procedure presenting a risk of serious injury to the patient.

The adjustable-length rod is intended for *in situ* adjustment after placement of the hooks or screws during spinal fusion surgery and is intended for use as before either a single or double rod assembly. It allows for distraction at a central location once bone anchors have been secured.

The bullet end rods are intended for use with the Jackson Intracanal Fixation Technique.

Disclaimer: This document is intended exclusively for physicians and is not intended for laypersons. Information on the products and procedures contained in this document is of a general nature and does not represent and does not constitute medical advice or recommendations. Because this information does not purport to constitute any diagnostic or therapeutic statement with regard to any individual medical case, each patient must be examined and advised individually, and this document does not replace the need for such examination and/or advice in whole or in part.



Caution: Federal (USA) law restricts this device to sale by or on the order of a physician. Rx Only. Please refer to the package inserts for important product information, including, but not limited to, indications, contraindications, warnings, precautions, adverse effects, and patient counseling information.

eLabeling: The Instructions for Use can be accessed online by visiting the website and using the KEY-CODE provided on the product label and as shown below. Additional translations are also available in electronic format for download. To request a paper copy of the Instructions for Use, contact Zimmer Biomet Spine at the phone number provided.



Consult Instructions for Use on this website:

<http://labeling.zimmerbiomet.com>

Key-Code: 060505-02



Manufactured by:

Zimmer Biomet Spine, Inc.
10225 Westmoor Dr.
Westminster, CO 80021 USA
+1 800.447.3625



Zimmer GmbH
Sulzerallee 8
CH-8404 Winterthur
Switzerland
+41 058.854.80.00



ZIMMER BIOMET

Your progress. Our promise.®

800.447.3625/zimmerbiomet.com

©2020 Zimmer Biomet Spine, Inc. All rights reserved.

All content herein is protected by copyright, trademarks and other intellectual property rights, as applicable, owned by or licensed to Zimmer Biomet Spine, Inc. or its affiliates unless otherwise indicated, and must not be redistributed, duplicated or disclosed, in whole or in part, without the express written consent of Zimmer Biomet Spine. This material is intended for health care professionals, the Zimmer Biomet Spine sales force and authorized representatives. Distribution to any other recipient is prohibited. Crossbar is a trademark of SeaSpine, Inc. Helical Flange is a registered trademark of Roger P. Jackson.

0744.1-US-en-REV1120